

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034704.D
 Acq On : 23 Mar 2023 17:25
 Operator : JC/MD
 Sample : 02055-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 24 05:47:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	316084	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	530117	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	464553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	204835	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	222435	48.372	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.740%
35) Dibromofluoromethane	5.391	113	163590	46.739	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.480%
50) Toluene-d8	8.646	98	631726	48.349	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.700%
62) 4-Bromofluorobenzene	11.079	95	239521	45.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.760%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1029	0.245	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034704.D
Acq On : 23 Mar 2023 17:25
Operator : JC/MD
Sample : 02055-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 24 05:47:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

