

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
 Data File : VX034707.D
 Acq On : 23 Mar 2023 18:35
 Operator : JC/MD
 Sample : 02055-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Mar 24 05:48:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	299863	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	507798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	445285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	220292	50.498	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.000%			
35) Dibromofluoromethane	5.385	113	160196	47.781	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.560%			
50) Toluene-d8	8.647	98	603043	48.182	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 96.360%			
62) 4-Bromofluorobenzene	11.079	95	226876	44.877	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.760%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	3431	1.670	ug/l	94
20) Methylene Chloride	2.788	84	1335	0.335	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034707.D
Acq On : 23 Mar 2023 18:35
Operator : JC/MD
Sample : 02055-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Mar 24 05:48:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

