


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032323\
Data File : VX034683.D
Acq On : 23 Mar 2023 08:32
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Title : SW846 8260
Last Update : Mon Mar 13 10:29:32 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	8253	PASS
75	95	30	60	53.6	22395	PASS
95	95	100	100	100.0	41749	PASS
96	95	5	9	6.3	2644	PASS
173	174	0.00	2	1.2	349	PASS
174	95	50	100	70.2	29315	PASS
175	174	5	9	7.0	2038	PASS
176	174	95	101	96.2	28213	PASS
177	176	5	9	6.3	1766	PASS