

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032618\
 Data File : VX000440.D
 Acq On : 26 Mar 2018 15:00
 Operator : JC/MD
 Sample : J1758-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-02-032318-C

Quant Time: Mar 27 07:49:24 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	110933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	193316	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	185577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	90199	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	82268	56.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.62%	
35) Dibromofluoromethane	5.52	113	63355	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
50) Toluene-d8	8.73	98	241550	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
62) 4-Bromofluorobenzene	11.15	95	78123	38.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.60%	
Target Compounds						
16) Acetone	2.45	43	8274	6.87	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032618\
Data File : VX000440.D
Acq On : 26 Mar 2018 15:00
Operator : JC/MD
Sample : J1758-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-02-032318-C

Quant Time: Mar 27 07:49:24 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

