

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032618\
 Data File : VX000442.D
 Acq On : 26 Mar 2018 15:51
 Operator : JC/MD
 Sample : J1760-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-032318-A

Quant Time: Mar 27 08:01:55 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	111284	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	200141	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	188670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	92851	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	82538	56.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.64%	
35) Dibromofluoromethane	5.52	113	63990	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
50) Toluene-d8	8.73	98	244361	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
62) 4-Bromofluorobenzene	11.15	95	80800	38.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.54%	
Target Compounds						
16) Acetone	2.45	43	2339	1.94	ug/l	90
20) Methylene Chloride	2.87	84	2175	1.56	ug/l #	79
95) Naphthalene	13.84	128	41618	5.49	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032618\
Data File : VX000442.D
Acq On : 26 Mar 2018 15:51
Operator : JC/MD
Sample : J1760-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-04-032318-A

Quant Time: Mar 27 08:01:55 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

