

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032618\
 Data File : VX000453.D
 Acq On : 26 Mar 2018 20:28
 Operator : JC/MD
 Sample : J2056-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22136-51-2-5-6-7-9

Manual Integrations
 APPROVED

sam
 3/27/2018 5:24:13 PM

Quant Time: Mar 27 09:49:01 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	111003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	201700	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	190327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	99769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	82902	57.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.42%	
35) Dibromofluoromethane	5.51	113	63800	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
50) Toluene-d8	8.73	98	250058	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
62) 4-Bromofluorobenzene	11.15	95	86130	40.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.24%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	15552	31.20	ug/l	97
16) Acetone	2.45	43	51277	42.54	ug/l	99
18) Methyl Acetate	2.79	43	26891	10.64	ug/l	99
25) 2-Butanone	4.71	43	52329	38.62	ug/l	99
40) Benzene	6.16	78	50099	9.13	ug/l	93
52) Toluene	8.79	92	15621	4.07	ug/l	97
68) m/p-Xylenes	10.37	106	3299m	1.17	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032618\
Data File : VX000453.D
Acq On : 26 Mar 2018 20:28
Operator : JC/MD
Sample : J2056-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22136-51-2-5-6-7-9

Manual Integrations
APPROVED

sam
3/27/2018 5:24:13 PM

Quant Time: Mar 27 09:49:01 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

