

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032718\
 Data File : VX000463.D
 Acq On : 27 Mar 2018 17:21
 Operator : JC/MD
 Sample : J2071-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1-MH-D-D

Quant Time: Mar 28 07:32:27 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	99376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	180919	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	175202	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	86948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	73367	56.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.12%	
35) Dibromofluoromethane	5.51	113	57306	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
50) Toluene-d8	8.73	98	224910	47.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.28%	
62) 4-Bromofluorobenzene	11.15	95	76058	39.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.18%	
Target Compounds						
16) Acetone	2.45	43	8278	7.67	ug/l	96
17) Carbon Disulfide	2.58	76	7264	2.55	ug/l	99
20) Methylene Chloride	2.86	84	6334	5.08	ug/l	94
68) m/p-Xylenes	10.37	106	3167	1.22	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	6295	1.24	ug/l	91
95) Naphthalene	13.84	128	123448	17.38	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032718\
Data File : VX000463.D
Acq On : 27 Mar 2018 17:21
Operator : JC/MD
Sample : J2071-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1-MH-D-D

Quant Time: Mar 28 07:32:27 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

