

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032718\
 Data File : VX000465.D
 Acq On : 27 Mar 2018 18:12
 Operator : JC/MD
 Sample : J2063-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RO-CV-2533

Quant Time: Mar 28 07:37:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	100615	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	183284	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	176041	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	85767	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	75445	57.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.88%	
35) Dibromofluoromethane	5.51	113	59772	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
50) Toluene-d8	8.73	98	230273	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
62) 4-Bromofluorobenzene	11.15	95	74666	38.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.12%	
Target Compounds						
16) Acetone	2.45	43	8787	8.04	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032718\
Data File : VX000465.D
Acq On : 27 Mar 2018 18:12
Operator : JC/MD
Sample : J2063-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RO-CV-2533

Quant Time: Mar 28 07:37:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

