

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032718\
 Data File : VX000474.D
 Acq On : 27 Mar 2018 22:01
 Operator : JC/MD
 Sample : J2062-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 19643

Quant Time: Mar 28 08:09:27 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	106973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	194108	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	194013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	112492	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	79828	57.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.34%	
35) Dibromofluoromethane	5.52	113	61556	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
50) Toluene-d8	8.73	98	243339	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
62) 4-Bromofluorobenzene	11.15	95	94790	44.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	48036	41.35	ug/l	95
19) Methyl tert-butyl Ether	3.22	73	4992	1.16	ug/l	94
25) 2-Butanone	4.72	43	8160	6.25	ug/l #	90
40) Benzene	6.15	78	179711	34.04	ug/l	98
52) Toluene	8.79	92	454479	123.12	ug/l	99
68) m/p-Xylenes	10.37	106	357788	124.23	ug/l	98
69) o-Xylene	10.71	106	282355	101.25	ug/l	96
80) 1,3,5-Trimethylbenzene	11.51	105	142494	22.74	ug/l	100
84) 1,2,4-Trimethylbenzene	11.82	105	628513	95.66	ug/l	100
95) Naphthalene	13.84	128	757424	82.40	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032718\
Data File : VX000474.D
Acq On : 27 Mar 2018 22:01
Operator : JC/MD
Sample : J2062-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
19643

Quant Time: Mar 28 08:09:27 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

