

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032818\
 Data File : VX000485.D
 Acq On : 28 Mar 2018 13:44
 Operator : JC/MD
 Sample : J1754-12 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-032718-A

Quant Time: Mar 29 04:29:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	101014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	180484	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	173017	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	85492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	73537	55.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.54%	
35) Dibromofluoromethane	5.52	113	58695	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
50) Toluene-d8	8.73	98	225849	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
62) 4-Bromofluorobenzene	11.15	95	76186	39.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.46%	
Target Compounds						
16) Acetone	2.45	43	9692	8.84	ug/l	95
20) Methylene Chloride	2.87	84	1415	1.12	ug/l #	80
95) Naphthalene	13.84	128	22123	3.17	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032818\
Data File : VX000485.D
Acq On : 28 Mar 2018 13:44
Operator : JC/MD
Sample : J1754-12 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-032718-A

Quant Time: Mar 29 04:29:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

