

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032818\
 Data File : VX000489.D
 Acq On : 28 Mar 2018 15:26
 Operator : JC/MD
 Sample : J1754-20 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-032718-E

Quant Time: Mar 29 04:37:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	95000	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	170260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	163163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	83615	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	70479	56.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.66%	
35) Dibromofluoromethane	5.52	113	53687	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
50) Toluene-d8	8.73	98	211669	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
62) 4-Bromofluorobenzene	11.15	95	71525	39.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.14%	
Target Compounds						
16) Acetone	2.45	43	9600	9.31	ug/l	96
20) Methylene Chloride	2.86	84	28814	24.18	ug/l	91
95) Naphthalene	13.84	128	8299	1.21	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032818\
Data File : VX000489.D
Acq On : 28 Mar 2018 15:26
Operator : JC/MD
Sample : J1754-20 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-032718-E

Quant Time: Mar 29 04:37:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

