

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032818\
 Data File : VX000491.D
 Acq On : 28 Mar 2018 16:17
 Operator : JC/MD
 Sample : J2086-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RS-124023-124035

Quant Time: Mar 29 04:41:04 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	94107	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	169589	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	166737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	84932	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	70204	57.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.30%	
35) Dibromofluoromethane	5.52	113	54661	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
50) Toluene-d8	8.73	98	213816	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
62) 4-Bromofluorobenzene	11.15	95	71725	39.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.58%	
Target Compounds						
16) Acetone	2.45	43	6709	6.57	ug/l	91
95) Naphthalene	13.84	128	50755	7.31	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032818\
Data File : VX000491.D
Acq On : 28 Mar 2018 16:17
Operator : JC/MD
Sample : J2086-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RS-124023-124035

Quant Time: Mar 29 04:41:04 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

