

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
 Data File : VX000514.D
 Acq On : 29 Mar 2018 15:58
 Operator : JC/MD
 Sample : J2062-05DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 19626-7-8-74DL

Quant Time: Mar 30 07:05:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	131859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	240400	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	250021	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	139568	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	99120	57.59	ug/l	0.00
Spiked Amount			Recovery	=	115.18%	
35) Dibromofluoromethane	5.51	113	74496	49.18	ug/l	0.00
Spiked Amount			Recovery	=	98.36%	
50) Toluene-d8	8.73	98	309380	49.32	ug/l	0.00
Spiked Amount			Recovery	=	98.64%	
62) 4-Bromofluorobenzene	11.15	95	119584	45.71	ug/l	0.00
Spiked Amount			Recovery	=	91.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.08	59	6333	10.70	ug/l #	83
16) Acetone	2.45	43	25839	18.05	ug/l	99
18) Methyl Acetate	2.79	43	30664	10.22	ug/l	96
25) 2-Butanone	4.73	43	4979	3.09	ug/l #	79
40) Benzene	6.15	78	233419	35.69	ug/l	99
52) Toluene	8.79	92	183860	40.22	ug/l	98
67) Ethyl Benzene	10.26	91	30102	3.19	ug/l	99
68) m/p-Xylenes	10.37	106	79471	21.41	ug/l	98
69) o-Xylene	10.71	106	69447	19.32	ug/l	99
70) Styrene	10.72	104	38533	6.48	ug/l #	85
80) 1,3,5-Trimethylbenzene	11.51	105	20692	2.66	ug/l	99
84) 1,2,4-Trimethylbenzene	11.82	105	77243	9.48	ug/l	98
95) Naphthalene	13.84	128	1248134	109.45	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
Data File : VX000514.D
Acq On : 29 Mar 2018 15:58
Operator : JC/MD
Sample : J2062-05DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
19626-7-8-74DL

Quant Time: Mar 30 07:05:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

