

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
 Data File : VX000515.D
 Acq On : 29 Mar 2018 16:24
 Operator : JC/MD
 Sample : J2097-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-4

Quant Time: Mar 30 07:10:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	122770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	228523	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	224699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	118300	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	94023	58.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.34%	
35) Dibromofluoromethane	5.51	113	70665	49.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.16%	
50) Toluene-d8	8.73	98	286243	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	
62) 4-Bromofluorobenzene	11.15	95	102533	41.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.68%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
Data File : VX000515.D
Acq On : 29 Mar 2018 16:24
Operator : JC/MD
Sample : J2097-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-4

Quant Time: Mar 30 07:10:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

