

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
 Data File : VX000522.D
 Acq On : 29 Mar 2018 19:23
 Operator : JC/MD
 Sample : J2101-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-032718

Quant Time: Mar 30 07:27:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	115254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	203702	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	211048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	110933	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	84072	55.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.76%	
35) Dibromofluoromethane	5.51	113	65191	50.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.58%	
50) Toluene-d8	8.73	98	266580	50.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.30%	
62) 4-Bromofluorobenzene	11.15	95	94970	43.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
Data File : VX000522.D
Acq On : 29 Mar 2018 19:23
Operator : JC/MD
Sample : J2101-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-032718

Quant Time: Mar 30 07:27:31 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

