

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
 Data File : VX000526.D
 Acq On : 29 Mar 2018 21:05
 Operator : JC/MD
 Sample : J2098-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LIC-PY-TRIP-BLANK-3.27.18

Quant Time: Mar 30 07:52:20 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	111355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	195006	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	203375	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108133	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	78375	53.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.84%	
35) Dibromofluoromethane	5.51	113	61222	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
50) Toluene-d8	8.73	98	257959	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	11.15	95	92019	43.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.36%	
Target Compounds						
16) Acetone	2.45	43	7555	6.25	ug/l	97
20) Methylene Chloride	2.86	84	2264	1.62	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
Data File : VX000526.D
Acq On : 29 Mar 2018 21:05
Operator : JC/MD
Sample : J2098-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LIC-PY-TRIP-BLANK-3.27.18

Quant Time: Mar 30 07:52:20 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

