

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
 Data File : VX000536.D
 Acq On : 30 Mar 2018 02:11
 Operator : JC/MD
 Sample : J1761-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-04-032818

Quant Time: Mar 30 08:11:05 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	107306	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	189667	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	195544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	101539	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	77665	55.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.90%	
35) Dibromofluoromethane	5.52	113	60726	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
50) Toluene-d8	8.73	98	249063	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.15	95	86529	42.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.88%	
Target Compounds						
16) Acetone	2.45	43	7876	6.76	ug/l	99
20) Methylene Chloride	2.86	84	13311	9.89	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
Data File : VX000536.D
Acq On : 30 Mar 2018 02:11
Operator : JC/MD
Sample : J1761-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-04-032818

Quant Time: Mar 30 08:11:05 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

