

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
 Data File : VX000538.D
 Acq On : 30 Mar 2018 03:02
 Operator : JC/MD
 Sample : J2117-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TPIAS

Quant Time: Mar 30 08:15:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	105883	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	186761	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	190199	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	102698	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	76145	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
35) Dibromofluoromethane	5.51	113	59369	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	8.73	98	240628	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	11.15	95	84935	42.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.64%	
Target Compounds						
16) Acetone	2.45	43	8467	7.36	ug/l	90
20) Methylene Chloride	2.87	84	8357	6.29	ug/l	97
95) Naphthalene	13.84	128	70046	8.35	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX032918\
Data File : VX000538.D
Acq On : 30 Mar 2018 03:02
Operator : JC/MD
Sample : J2117-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TPIAS

Quant Time: Mar 30 08:15:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

