

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032923\
 Data File : VX034826.D
 Acq On : 29 Mar 2023 14:13
 Operator : JC/MD
 Sample : VX0329WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0329WBL01

Quant Time: Apr 01 02:05:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	271992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	462991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	410636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	188645	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	191181	45.960	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.920%
35) Dibromofluoromethane	5.379	113	145980	47.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.340%
50) Toluene-d8	8.646	98	556338	49.220	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.079	95	204212	46.687	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.380%
Target Compounds						
					Qvalue	
89) n-Butylbenzene	12.335	91	4970	0.517	ug/l	93
90) Hexachloroethane	12.542	117	402	0.236	ug/l	67
93) 1,2,4-Trichlorobenzene	13.591	180	1584	0.448	ug/l	88
94) Hexachlorobutadiene	13.725	225	2854	1.837	ug/l	97
95) Naphthalene	13.780	128	3099	0.247	ug/l	98
96) 1,2,3-Trichlorobenzene	13.963	180	1861	0.522	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032923\
Data File : VX034826.D
Acq On : 29 Mar 2023 14:13
Operator : JC/MD
Sample : VX0329WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0329WBL01

Quant Time: Apr 01 02:05:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

