

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032923\
 Data File : VX034836.D
 Acq On : 29 Mar 2023 18:18
 Operator : JC/MD
 Sample : 02084-03DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6ADL

Quant Time: Mar 30 06:33:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 30 05:38:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	247217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	423017	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174378	46.122	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.240%
35) Dibromofluoromethane	5.391	113	131056	46.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	8.647	98	512108	49.588	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.180%
62) 4-Bromofluorobenzene	11.079	95	189439	47.402	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.800%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4557	2.967	ug/l	96
31) Cyclohexane	5.477	56	35339	7.676	ug/l	95
39) Methylcyclohexane	7.379	83	21590	4.905	ug/l	88
40) Benzene	6.037	78	594639	49.973	ug/l	100
52) Toluene	8.720	92	2470750	318.457	ug/l	97
67) Ethyl Benzene	10.195	91	515935	36.641	ug/l	98
68) m/p-Xylenes	10.299	106	661492	122.982	ug/l	98
69) o-Xylene	10.640	106	240274	45.599	ug/l	98
73) Isopropylbenzene	10.963	105	19810	1.567	ug/l	100
78) n-propylbenzene	11.305	91	63831	4.370	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	133710	12.647	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	432528	40.229	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032923\
Data File : VX034836.D
Acq On : 29 Mar 2023 18:18
Operator : JC/MD
Sample : 02084-03DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6ADL

Quant Time: Mar 30 06:33:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 30 05:38:28 2023
Response via : Initial Calibration

