



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032923\
Data File : VX034816.D
Acq On : 29 Mar 2023 08:36
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Title : SW846 8260
Last Update : Thu Mar 30 11:02:25 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	13738	PASS
75	95	30	60	53.4	35456	PASS
95	95	100	100	100.0	66429	PASS
96	95	5	9	6.9	4584	PASS
173	174	0.00	2	1.1	529	PASS
174	95	50	100	72.2	47971	PASS
175	174	5	9	7.3	3510	PASS
176	174	95	101	95.7	45915	PASS
177	176	5	9	7.0	3204	PASS