

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
 Data File : VX000560.D
 Acq On : 30 Mar 2018 14:54
 Operator : JC/MD
 Sample : J2139-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 03 17:09:10 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	96989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	167735	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	168621	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	87670	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	69805	55.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.28%	
35) Dibromofluoromethane	5.52	113	54405	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
50) Toluene-d8	8.73	98	213954	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
62) 4-Bromofluorobenzene	11.15	95	72992	40.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.52%	
Target Compounds						
16) Acetone	2.45	43	3457	3.282	ug/l #	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
Data File : VX000560.D
Acq On : 30 Mar 2018 14:54
Operator : JC/MD
Sample : J2139-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 03 17:09:10 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

