

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
 Data File : VX000567.D
 Acq On : 30 Mar 2018 18:22
 Operator : JC/MD
 Sample : J2106-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 19622

Manual Integrations
 APPROVED

sam
 4/4/2018 6:11:16 PM

Quant Time: Apr 04 10:53:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	160148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	287596	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	362776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	154078	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	107199	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
35) Dibromofluoromethane	5.52	113	87509	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
50) Toluene-d8	8.73	98	371054	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.15	95	141130	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
						Qvalue
13) Acrolein	2.29	56	18997	92.49	ug/l	98
16) Acetone	2.45	43	1257200	722.93	ug/l	99
18) Methyl Acetate	2.79	43	732556	200.93	ug/l	98
20) Methylene Chloride	2.87	84	50319	25.05	ug/l	93
25) 2-Butanone	4.72	43	19006	9.72	ug/l	91
30) Chloroform	5.23	83	81422	25.46	ug/l	100
40) Benzene	6.15	78	21166	2.71	ug/l	98
44) Trichloroethene	7.22	130	10520	4.69	ug/l	99
47) Bromodichloromethane	7.92	83	6072	2.20	ug/l	93
52) Toluene	8.79	92	25435	4.65	ug/l	98
67) Ethyl Benzene	10.26	91	11836	0.86	ug/l	90
68) m/p-Xylenes	10.37	106	12384	2.30	ug/l	98
69) o-Xylene	10.71	106	9101	1.75	ug/l	94
76) 1,2,3-Trichloropropane	11.30	75	3147	0.95	ug/l #	36
84) 1,2,4-Trimethylbenzene	11.82	105	10694	1.19	ug/l	77
86) p-Isopropyltoluene	12.07	119	21189	2.23	ug/l	89
95) Naphthalene	13.84	128	16382m	1.30	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
Data File : VX000567.D
Acq On : 30 Mar 2018 18:22
Operator : JC/MD
Sample : J2106-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
19622

Manual Integrations
APPROVED

sam
4/4/2018 6:11:16 PM

Quant Time: Apr 04 10:53:53 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

