

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
 Data File : VX000569.D
 Acq On : 30 Mar 2018 19:14
 Operator : JC/MD
 Sample : J2131-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 4534-37

Quant Time: Apr 04 11:16:56 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	150843	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	266915	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	274781	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144464	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	107973	54.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.68%	
35) Dibromofluoromethane	5.51	113	82434	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
50) Toluene-d8	8.73	98	340289	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	11.15	95	128718	44.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.00%	
Target Compounds						
16) Acetone	2.45	43	13291	8.11	ug/l	98
20) Methylene Chloride	2.87	84	27806	14.70	ug/l	90
95) Naphthalene	13.84	128	17197	1.46	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
Data File : VX000569.D
Acq On : 30 Mar 2018 19:14
Operator : JC/MD
Sample : J2131-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
4534-37

Quant Time: Apr 04 11:16:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

