

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
 Data File : VX000574.D
 Acq On : 30 Mar 2018 21:21
 Operator : JC/MD
 Sample : J2131-05 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHRT24256

Quant Time: Apr 04 11:23:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	132301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	232255	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	231609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	121972	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	95338	55.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.40%	
35) Dibromofluoromethane	5.51	113	72532	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
50) Toluene-d8	8.73	98	293254	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
62) 4-Bromofluorobenzene	11.15	95	107023	42.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.60%	
Target Compounds						
16) Acetone	2.45	43	9023	6.28	ug/l	95
20) Methylene Chloride	2.87	84	27238	16.41	ug/l	96
95) Naphthalene	13.84	128	22359	2.24	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
Data File : VX000574.D
Acq On : 30 Mar 2018 21:21
Operator : JC/MD
Sample : J2131-05 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHRT24256

Quant Time: Apr 04 11:23:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

