

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
 Data File : VX000576.D
 Acq On : 30 Mar 2018 22:12
 Operator : JC/MD
 Sample : J2131-07 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHRT24482

Quant Time: Mar 31 05:55:24 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	129891	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	221415	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	223280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	120529	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	90357	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
35) Dibromofluoromethane	5.51	113	69845	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
50) Toluene-d8	8.73	98	281080	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.15	95	102525	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
Target Compounds						
16) Acetone	2.45	43	7521	5.332	ug/l	93
20) Methylene Chloride	2.87	84	105439	64.721	ug/l	97
95) Naphthalene	13.84	128	16796	1.705	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX033018\
Data File : VX000576.D
Acq On : 30 Mar 2018 22:12
Operator : JC/MD
Sample : J2131-07 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHRT24482

Quant Time: Mar 31 05:55:24 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

