

Data File : VX000591.D
Acq On : 31 Mar 2018 14:28
Operator : JC/MD
Sample : J2142-08 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GB-RT-4252

Quant Time: Apr 02 06:25:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	128255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	218183	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	216692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	114228	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	86654	51.76	ug/l	0.00
Spiked Amount			Recovery	=	103.52%	
35) Dibromofluoromethane	5.52	113	65620	47.73	ug/l	0.00
Spiked Amount			Recovery	=	95.46%	
50) Toluene-d8	8.73	98	271414	47.67	ug/l	0.00
Spiked Amount			Recovery	=	95.34%	
62) 4-Bromofluorobenzene	11.15	95	97502	41.70	ug/l	0.00
Spiked Amount			Recovery	=	83.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	6347	4.56	ug/l	93
95) Naphthalene	13.84	128	22746	2.44	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000591.D
Acq On : 31 Mar 2018 14:28
Operator : JC/MD
Sample : J2142-08 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GB-RT-4252

Quant Time: Apr 02 06:25:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

