

Data File : VX000593.D
Acq On : 31 Mar 2018 15:17
Operator : JC/MD
Sample : J1752-04 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-01-032918-A

Quant Time: Apr 02 07:06:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	124282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	209984	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	204426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	111278	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	82559	50.89	ug/l	0.00
Spiked Amount			Recovery	=	101.78%	
35) Dibromofluoromethane	5.51	113	65180	49.26	ug/l	0.00
Spiked Amount			Recovery	=	98.52%	
50) Toluene-d8	8.73	98	263379	48.06	ug/l	0.00
Spiked Amount			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	11.15	95	94121	41.80	ug/l	0.00
Spiked Amount			Recovery	=	83.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	7590	5.62	ug/l	96
20) Methylene Chloride	2.86	84	3073	1.97	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000593.D
Acq On : 31 Mar 2018 15:17
Operator : JC/MD
Sample : J1752-04 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-01-032918-A

Quant Time: Apr 02 07:06:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

