

Data File : VX000594.D
Acq On : 31 Mar 2018 15:43
Operator : JC/MD
Sample : J1752-06 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-01-032918-B

Quant Time: Apr 02 07:08:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	125929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	212999	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	212378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113297	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	85675	52.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.24%	
35) Dibromofluoromethane	5.51	113	65935	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
50) Toluene-d8	8.73	98	269448	48.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.96%	
62) 4-Bromofluorobenzene	11.15	95	95852	41.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.88%	

Target Compounds				Qvalue		
16) Acetone	2.45	43	8632	6.31	ug/l	93
20) Methylene Chloride	2.86	84	101153	64.04	ug/l	97
95) Naphthalene	13.84	128	153882	16.62	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000594.D
Acq On : 31 Mar 2018 15:43
Operator : JC/MD
Sample : J1752-06 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-01-032918-B

Quant Time: Apr 02 07:08:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

