

Data File : VX000596.D
Acq On : 31 Mar 2018 16:34
Operator : JC/MD
Sample : J1760-10 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-03-032918

Quant Time: Apr 04 04:16:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	120812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	202468	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	199510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	105784	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	80937	51.32	ug/l	0.00
Spiked Amount			Recovery	=	102.64%	
35) Dibromofluoromethane	5.52	113	62182	48.74	ug/l	0.00
Spiked Amount			Recovery	=	97.48%	
50) Toluene-d8	8.72	98	256522	48.55	ug/l	0.00
Spiked Amount			Recovery	=	97.10%	
62) 4-Bromofluorobenzene	11.15	95	90574	41.73	ug/l	0.00
Spiked Amount			Recovery	=	83.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	7611	5.80	ug/l #	85
20) Methylene Chloride	2.86	84	1900	1.25	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000596.D
Acq On : 31 Mar 2018 16:34
Operator : JC/MD
Sample : J1760-10 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-03-032918

Quant Time: Apr 04 04:16:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

