

Data File : VX000597.D
Acq On : 31 Mar 2018 16:59
Operator : JC/MD
Sample : J2146-02 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-CS

Quant Time: Apr 02 07:16:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	115314	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	197434	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	194753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	101339	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	78924	52.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.86%	
35) Dibromofluoromethane	5.51	113	60604	48.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.44%	
50) Toluene-d8	8.73	98	245245	47.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.20%	
62) 4-Bromofluorobenzene	11.15	95	85971	40.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.58%	

Target Compounds				Qvalue		
16) Acetone	2.45	43	7570	6.05	ug/l #	86
20) Methylene Chloride	2.86	84	82832	57.27	ug/l	97
95) Naphthalene	13.84	128	12963	1.57	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000597.D
Acq On : 31 Mar 2018 16:59
Operator : JC/MD
Sample : J2146-02 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-CS

Quant Time: Apr 02 07:16:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

