

Data File : VX000598.D
Acq On : 31 Mar 2018 17:25
Operator : JC/MD
Sample : J2146-04 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-CD

Quant Time: Apr 04 04:39:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	119901	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	203264	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	196897	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	105756	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	81716	52.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.42%	
35) Dibromofluoromethane	5.51	113	64256	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
50) Toluene-d8	8.73	98	254274	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	
62) 4-Bromofluorobenzene	11.15	95	89309	41.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.20%	

Target Compounds				Qvalue		
16) Acetone	2.45	43	8328	6.40	ug/l	98
20) Methylene Chloride	2.86	84	15470729	10287.59	ug/l	86
28) Bromochloromethane	5.04	49	23211	22.04	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000598.D
Acq On : 31 Mar 2018 17:25
Operator : JC/MD
Sample : J2146-04 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-CD

Quant Time: Apr 04 04:39:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

