

Data File : VX000601.D
Acq On : 31 Mar 2018 18:42
Operator : JC/MD
Sample : J2150-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Apr 03 19:01:12 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	108626	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	183579	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	182780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	99986	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	73693	51.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.94%	
35) Dibromofluoromethane	5.51	113	59199	51.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.36%	
50) Toluene-d8	8.73	98	230154	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
62) 4-Bromofluorobenzene	11.15	95	82302	41.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.62%	

Target Compounds						Qvalue
16) Acetone	2.45	43	8992	7.62	ug/l	96
19) Methyl tert-butyl Ether	3.21	73	191277	43.93	ug/l	98
29) Tetrahydrofuran	5.19	42	6511	7.98	ug/l	94
40) Benzene	6.16	78	3026	0.61	ug/l #	68
52) Toluene	8.79	92	4708	1.35	ug/l	92
67) Ethyl Benzene	10.26	91	6983	1.01	ug/l	92
68) m/p-Xylenes	10.36	106	5289	1.95	ug/l	89
69) o-Xylene	10.71	106	2515	0.96	ug/l	96
70) Styrene	10.72	104	3304	0.76	ug/l	95
80) 1,3,5-Trimethylbenzene	11.51	105	1942	0.35	ug/l	90
84) 1,2,4-Trimethylbenzene	11.82	105	6223	1.07	ug/l	90
95) Naphthalene	13.84	128	5735	0.70	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000601.D

Acq On : 31 Mar 2018 18:42

Operator : JC/MD

Sample : J2150-08

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 20 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-8

Quant Time: Apr 03 19:01:12 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M

Quant Title : SW846 8260

QLast Update : Wed Mar 21 04:52:42 2018

Response via : Initial Calibration

