

Data File : VX000604.D
Acq On : 31 Mar 2018 19:59
Operator : JC/MD
Sample : J2150-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Apr 04 04:58:45 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	115404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	192740	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	189613	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	103226	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	77810	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
35) Dibromofluoromethane	5.51	113	60896	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	8.73	98	239035	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
62) 4-Bromofluorobenzene	11.15	95	86113	41.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
19) Methyl tert-butyl Ether	3.21	73	28262	6.11	ug/l	97
29) Tetrahydrofuran	5.18	42	12629	14.57	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000604.D

Acq On : 31 Mar 2018 19:59

Operator : JC/MD

Sample : J2150-04

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 23 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-4

Quant Time: Apr 04 04:58:45 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M

Quant Title : SW846 8260

QLast Update : Wed Mar 21 04:52:42 2018

Response via : Initial Calibration

