

Data File : VX000613.D
Acq On : 01 Apr 2018 00:39
Operator : JC/MD
Sample : J2132-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 02 07:04:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	109841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	180921	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	181409	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95971	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	71489	49.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.72%	
35) Dibromofluoromethane	5.52	113	56255	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	
50) Toluene-d8	8.73	98	228992	48.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.00%	
62) 4-Bromofluorobenzene	11.15	95	79677	41.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.36%	

Target Compounds						Qvalue
16) Acetone	2.45	43	3166	2.65	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000613.D
Acq On : 01 Apr 2018 00:39
Operator : JC/MD
Sample : J2132-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 02 07:04:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

