

Data File : VX000615.D
Acq On : 01 Apr 2018 01:30
Operator : JC/MD
Sample : J2150-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Apr 04 05:47:25 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	106557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	181229	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	175749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	96025	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	72012	51.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.54%	
35) Dibromofluoromethane	5.51	113	54685	47.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.78%	
50) Toluene-d8	8.73	98	231271	48.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.80%	
62) 4-Bromofluorobenzene	11.15	95	78349	40.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.08%	

Target Compounds				Qvalue		
11) Tert butyl alcohol	3.09	59	5072469	10601.83	ug/l	100
16) Acetone	2.45	43	5374	4.64	ug/l	98
19) Methyl tert-butyl Ether	3.21	73	3626894	849.07	ug/l	98
73) Isopropylbenzene	11.02	105	1724	0.27	ug/l	82
84) 1,2,4-Trimethylbenzene	11.82	105	1141	0.20	ug/l	99
85) sec-Butylbenzene	11.96	105	10491	1.57	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000615.D

Acq On : 01 Apr 2018 01:30

Operator : JC/MD

Sample : J2150-02

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 36 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-2R

Quant Time: Apr 04 05:47:25 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M

Quant Title : SW846 8260

QLast Update : Wed Mar 21 04:52:42 2018

Response via : Initial Calibration

