

Data File : VX000616.D
Acq On : 01 Apr 2018 01:56
Operator : JC/MD
Sample : J2150-02DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2RDL

Quant Time: Apr 02 07:17:08 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	109170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	182621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	177871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	95430	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.08	65	71687	50.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.60%	
35) Dibromofluoromethane	5.51	113	56382	49.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.00%	
50) Toluene-d8	8.72	98	227997	47.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.68%	
62) 4-Bromofluorobenzene	11.15	95	78303	40.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.52%	

Target Compounds

						Qvalue
11) Tert butyl alcohol	3.07	59	167902	342.53	ug/l	100
19) Methyl tert-butyl Ether	3.21	73	152857	34.93	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000616.D
Acq On : 01 Apr 2018 01:56
Operator : JC/MD
Sample : J2150-02DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2RDL

Quant Time: Apr 02 07:17:08 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

