

Data File : VX000617.D
Acq On : 01 Apr 2018 02:21
Operator : JC/MD
Sample : J2150-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Apr 04 05:59:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	111322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	185969	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	188109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	101683	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	76066	52.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.70%	
35) Dibromofluoromethane	5.51	113	58954	50.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.62%	
50) Toluene-d8	8.73	98	237109	48.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.72%	
62) 4-Bromofluorobenzene	11.15	95	82650	41.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.00%	

Target Compounds				Qvalue		
11) Tert butyl alcohol	3.07	59	460502	921.28	ug/l	100
17) Carbon Disulfide	2.57	76	4759	1.50	ug/l	97
19) Methyl tert-butyl Ether	3.21	73	3534834	792.09	ug/l	99
86) p-Isopropyltoluene	12.07	119	8067	1.28	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000617.D

Acq On : 01 Apr 2018 02:21

Operator : JC/MD

Sample : J2150-07

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 38 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-7

Quant Time: Apr 04 05:59:47 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M

Quant Title : SW846 8260

QLast Update : Wed Mar 21 04:52:42 2018

Response via : Initial Calibration

