

Data File : VX000618.D
Acq On : 01 Apr 2018 02:47
Operator : JC/MD
Sample : J2150-07DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7DL

Quant Time: Apr 02 07:21:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	107314	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	178756	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	178520	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95524	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	70947	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
35) Dibromofluoromethane	5.51	113	55843	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	8.73	98	222957	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
62) 4-Bromofluorobenzene	11.15	95	77383	40.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.07	59	68183	141.50	ug/l	99
19) Methyl tert-butyl Ether	3.21	73	627163	145.78	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000618.D
Acq On : 01 Apr 2018 02:47
Operator : JC/MD
Sample : J2150-07DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7DL

Quant Time: Apr 02 07:21:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

