

Data File : VX000619.D
Acq On : 01 Apr 2018 03:12
Operator : JC/MD
Sample : J2132-04 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW11

Quant Time: Apr 04 06:06:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	108862	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	177837	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	179232	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	103638	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	71517	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
35) Dibromofluoromethane	5.51	113	55944	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
50) Toluene-d8	8.72	98	230333	49.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.26%	
62) 4-Bromofluorobenzene	11.15	95	83176	43.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.70%	

Target Compounds					Qvalue	
31) Cyclohexane	5.59	56	18898	10.54	ug/l	97
39) Methylcyclohexane	7.48	83	9282	4.58	ug/l #	89
67) Ethyl Benzene	10.26	91	114489	16.93	ug/l	99
68) m/p-Xylenes	10.37	106	72169	27.12	ug/l	99
69) o-Xylene	10.71	106	18009	6.99	ug/l	98
73) Isopropylbenzene	11.02	105	9962	1.45	ug/l	95
78) n-propylbenzene	11.37	91	31816	3.89	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	36720	6.36	ug/l	100
84) 1,2,4-Trimethylbenzene	11.82	105	178329	29.46	ug/l	99
95) Naphthalene	13.84	128	59357	7.01	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000619.D

Acq On : 01 Apr 2018 03:12

Operator : JC/MD

Sample : J2132-04 20X

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 40 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW11

Quant Time: Apr 04 06:06:48 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M

Quant Title : SW846 8260

QLast Update : Wed Mar 21 04:52:42 2018

Response via : Initial Calibration

