

Data File : VX000620.D
Acq On : 01 Apr 2018 03:38
Operator : JC/MD
Sample : J2132-05 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW12

Quant Time: Apr 02 07:30:24 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	112145	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	182642	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	184068	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	107495	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	70213	47.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.92%	
35) Dibromofluoromethane	5.52	113	55367	48.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.22%	
50) Toluene-d8	8.73	98	228482	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	
62) 4-Bromofluorobenzene	11.15	95	86995	44.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.06%	

Target Compounds					Qvalue	
31) Cyclohexane	5.59	56	13344	7.23	ug/l	96
39) Methylcyclohexane	7.47	83	6091	2.93	ug/l	90
40) Benzene	6.15	78	6385	1.29	ug/l	100
52) Toluene	8.79	92	113110	32.57	ug/l	97
67) Ethyl Benzene	10.26	91	269735	38.84	ug/l	99
68) m/p-Xylenes	10.37	106	334700	122.49	ug/l	98
69) o-Xylene	10.71	106	80718	30.51	ug/l	97
73) Isopropylbenzene	11.02	105	8968	1.26	ug/l	99
78) n-propylbenzene	11.37	91	26257	3.09	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	42668	7.12	ug/l	98
84) 1,2,4-Trimethylbenzene	11.82	105	162271	25.84	ug/l	99
95) Naphthalene	13.84	128	94804	10.79	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
ALS Vial      : 41      Sample Multiplier: 1
```

MW12

Response via : Initial Calibration

