

Data File : VX000621.D
Acq On : 01 Apr 2018 04:03
Operator : JC/MD
Sample : J2132-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NW3

Quant Time: Apr 02 07:35:04 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	106919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	182533	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	180344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	102529	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71402	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
35) Dibromofluoromethane	5.51	113	55391	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
50) Toluene-d8	8.73	98	230356	48.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.72%	
62) 4-Bromofluorobenzene	11.15	95	82305	42.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.02%	

Target Compounds				Qvalue		
31) Cyclohexane	5.57	56	5893	3.35	ug/l	96
39) Methylcyclohexane	7.47	83	3178	1.53	ug/l	93
52) Toluene	8.79	92	3735	1.08	ug/l	87
67) Ethyl Benzene	10.26	91	82963	12.19	ug/l	100
68) m/p-Xylenes	10.37	106	85816	32.05	ug/l	97
69) o-Xylene	10.71	106	23595	9.10	ug/l	98
73) Isopropylbenzene	11.02	105	7145	1.05	ug/l	97
78) n-propylbenzene	11.37	91	21494	2.65	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	32685	5.72	ug/l	99
84) 1,2,4-Trimethylbenzene	11.82	105	132395	22.11	ug/l	98
95) Naphthalene	13.84	128	48659	5.81	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000621.D
Acq On : 01 Apr 2018 04:03
Operator : JC/MD
Sample : J2132-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NW3

Quant Time: Apr 02 07:35:04 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

