

Data File : VX000626.D
Acq On : 01 Apr 2018 06:10
Operator : JC/MD
Sample : J2132-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW8DL

Quant Time: Apr 04 07:19:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	121847	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	201515	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	198942	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	109947	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	80525	50.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.26%	
35) Dibromofluoromethane	5.52	113	61159	48.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.34%	
50) Toluene-d8	8.73	98	252956	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
62) 4-Bromofluorobenzene	11.15	95	93572	43.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.16%	

Target Compounds						Qvalue
31) Cyclohexane	5.58	56	13826	6.89	ug/l	# 79
39) Methylcyclohexane	7.47	83	6341	2.76	ug/l	94
40) Benzene	6.15	78	4230	0.77	ug/l	# 90
52) Toluene	8.79	92	22139	5.78	ug/l	94
67) Ethyl Benzene	10.26	91	184515	24.59	ug/l	99
68) m/p-Xylenes	10.37	106	99091	33.55	ug/l	99
69) o-Xylene	10.71	106	48767	17.05	ug/l	97
73) Isopropylbenzene	11.02	105	14441	1.98	ug/l	98
78) n-propylbenzene	11.37	91	38100	4.39	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	22201	3.62	ug/l	97
84) 1,2,4-Trimethylbenzene	11.82	105	152293	23.71	ug/l	99
95) Naphthalene	13.84	128	56090	6.24	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000626.D
Acq On : 01 Apr 2018 06:10
Operator : JC/MD
Sample : J2132-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW8DL

Quant Time: Apr 04 07:19:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

