

Data File : VX000630.D
Acq On : 01 Apr 2018 07:53
Operator : JC/MD
Sample : J2132-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NW4DL

Quant Time: Apr 02 08:14:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	122296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	201717	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	209533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128749	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	81800	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	
35) Dibromofluoromethane	5.51	113	61434	48.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.68%	
50) Toluene-d8	8.73	98	260895	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
62) 4-Bromofluorobenzene	11.15	95	100416	45.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.48%	

Target Compounds					Qvalue	
16) Acetone	2.47	43	3184	2.40	ug/l	89
39) Methylcyclohexane	7.47	83	12052	5.24	ug/l	98
67) Ethyl Benzene	10.26	91	6482	0.82	ug/l	94
68) m/p-Xylenes	10.37	106	2360	0.76	ug/l	100
73) Isopropylbenzene	11.02	105	6525	0.76	ug/l	99
78) n-propylbenzene	11.37	91	52899	5.20	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	37687	5.25	ug/l	98
84) 1,2,4-Trimethylbenzene	11.82	105	65375	8.69	ug/l	99
85) sec-Butylbenzene	11.95	105	8704	0.97	ug/l	89
95) Naphthalene	13.84	128	8494	0.81	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000630.D
Acq On : 01 Apr 2018 07:53
Operator : JC/MD
Sample : J2132-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NW4DL

Quant Time: Apr 02 08:14:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

