

Data Path : W:\HPCHEM1\MSVOA X\Data\VX040218\
 Data File : VX000642.D
 Acq On : 02 Apr 2018 15:50
 Operator : JC/MD
 Sample : J1757-08 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-03-033018-A

Quant Time: Apr 03 02:59:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 21 04:52:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	120793	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	193117	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	198033	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106005	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	77699	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
35) Dibromofluoromethane	5.51	113	60759	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
50) Toluene-d8	8.72	98	247826	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.15	95	89497	43.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.02%	
Target Compounds						
16) Acetone	2.45	43	8178	6.23	ug/l	# 87
20) Methylene Chloride	2.87	84	3966	2.62	ug/l	94
95) Naphthalene	13.84	128	77292	8.92	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\Data\VX040218\
Data File : VX000642.D
Acq On : 02 Apr 2018 15:50
Operator : JC/MD
Sample : J1757-08 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EO-03-033018-A

Quant Time: Apr 03 02:59:31 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X032018W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 21 04:52:42 2018
Response via : Initial Calibration

