

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
 Data File : VX000674.D
 Acq On : 04 Apr 2018 22:59
 Operator : JC/MD
 Sample : J2207-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-1

Quant Time: Apr 05 05:47:10 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 05 03:34:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	86536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	149286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	146630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	74625	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.09	65	63861	43.44	ug/l	0.00
Spiked Amount			Recovery	=	86.88%	
35) Dibromofluoromethane	5.51	113	47420	46.53	ug/l	0.00
Spiked Amount			Recovery	=	93.06%	
50) Toluene-d8	8.73	98	187782	43.33	ug/l	0.00
Spiked Amount			Recovery	=	86.66%	
62) 4-Bromofluorobenzene	11.15	95	65349	38.10	ug/l	0.00
Spiked Amount			Recovery	=	76.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	8193	6.77	ug/l	95
20) Methylene Chloride	2.86	84	174725	126.65	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
Data File : VX000674.D
Acq On : 04 Apr 2018 22:59
Operator : JC/MD
Sample : J2207-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-1

Quant Time: Apr 05 05:47:10 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 05 03:34:50 2018
Response via : Initial Calibration

