

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
 Data File : VX000675.D
 Acq On : 04 Apr 2018 23:25
 Operator : JC/MD
 Sample : J2151-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-06-040318-A

Quant Time: Apr 05 05:48:30 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 05 03:34:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	84286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	147368	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	146719	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	75718	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	67011	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
35) Dibromofluoromethane	5.51	113	47787	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
50) Toluene-d8	8.73	98	190859	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
62) 4-Bromofluorobenzene	11.15	95	66100	39.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.08%	
Target Compounds						
16) Acetone	2.45	43	7566	6.42	ug/l	91
20) Methylene Chloride	2.86	84	25296	18.83	ug/l	95
95) Naphthalene	13.84	128	19525	3.43	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
Data File : VX000675.D
Acq On : 04 Apr 2018 23:25
Operator : JC/MD
Sample : J2151-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
AU-06-040318-A

Quant Time: Apr 05 05:48:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 05 03:34:50 2018
Response via : Initial Calibration

