

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
 Data File : VX000676.D
 Acq On : 04 Apr 2018 23:50
 Operator : JC/MD
 Sample : J2151-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-06-040318-B

Quant Time: Apr 05 05:50:33 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 05 03:34:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	86229	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	145324	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	140502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	72345	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	64061	43.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.46%	
35) Dibromofluoromethane	5.51	113	46112	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
50) Toluene-d8	8.73	98	182282	43.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.42%	
62) 4-Bromofluorobenzene	11.15	95	64842	38.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.68%	
Target Compounds						
16) Acetone	2.45	43	6696	5.56	ug/l	95
20) Methylene Chloride	2.86	84	8274	6.02	ug/l	97
95) Naphthalene	13.84	128	9208	1.69	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
Data File : VX000676.D
Acq On : 04 Apr 2018 23:50
Operator : JC/MD
Sample : J2151-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-06-040318-B

Quant Time: Apr 05 05:50:33 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 05 03:34:50 2018
Response via : Initial Calibration

