

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
 Data File : VX000677.D
 Acq On : 05 Apr 2018 00:15
 Operator : JC/MD
 Sample : J2151-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-06-040318-C

Quant Time: Apr 05 05:52:00 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 05 03:34:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	83599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	143113	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	142553	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	73889	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	62265	43.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.68%	
35) Dibromofluoromethane	5.51	113	46455	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	8.73	98	181379	43.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.32%	
62) 4-Bromofluorobenzene	11.15	95	64414	39.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.36%	
Target Compounds						
16) Acetone	2.45	43	7256	6.21	ug/l	# 84
20) Methylene Chloride	2.86	84	1816	1.36	ug/l	91
95) Naphthalene	13.84	128	27037	4.86	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040418\
Data File : VX000677.D
Acq On : 05 Apr 2018 00:15
Operator : JC/MD
Sample : J2151-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-06-040318-C

Quant Time: Apr 05 05:52:00 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040418W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 05 03:34:50 2018
Response via : Initial Calibration

