

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
 Data File : VX000731.D
 Acq On : 08 Apr 2018 15:13
 Operator : JC/MD
 Sample : J2190-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18-040218

Quant Time: Apr 09 07:35:35 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 07 17:01:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	322813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	444435	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	414230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	243551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161886	43.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.70%	
35) Dibromofluoromethane	5.51	113	136279	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
50) Toluene-d8	8.72	98	513528	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
62) 4-Bromofluorobenzene	11.15	95	210298	46.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.02%	
Target Compounds						
16) Acetone	2.45	43	16440	14.20	ug/l	99
27) cis-1,2-Dichloroethene	4.61	96	4307	1.39	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
Data File : VX000731.D
Acq On : 08 Apr 2018 15:13
Operator : JC/MD
Sample : J2190-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18-040218

Quant Time: Apr 09 07:35:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 07 17:01:21 2018
Response via : Initial Calibration

